



Certificate of Analysis



Mesa Laboratories Inc
 12100 West 6th Ave., Lakewood, CO 80228
 TEL (303) 987-8000 FAX (303) 987-8989
www.mesalabs.com

Description: Conductivity Calibration Solution 50mS/cm @ 25°C
 Part Number: 02.0071
 Lot No: ML-C50-1443
 Certification Date: February 26, 2018
 Expiration Date: August 23, 2019

Nominal Value (mS/cm)	Measured Value (mS/cm)	Tolerance @ 25°C (mS/cm)	Measured Value In Tolerance	Reference System Standard	Reference System Uncertainty @ 25°C (mS/cm)
50.000	49.985	±0.4 mS	Yes	NIST SRM 999	±0.034

Test Methods

- All analytical balances are calibrated by an ISO/IEC 17025:2005 accredited calibration laboratory. All balances are checked prior to use using an in-house procedure. Weights used for testing are traceable to NIST.
- All thermometers are NIST traceable through reference temperature probes that are calibrated by an ISO/IEC 17025:2005 accredited calibration laboratory.
- Measurements are taken at 25°C ± 0.25°C using SOP 900215-001 and are temperature compensated.

Intended Use

This standard solution is indicated for calibrating conductivity meters.

Hazardous Information

Please refer to the Material Safety Data Sheet available on our website for information on this material

Stability and Storage

Protect from temperature extremes. Discard if solution has been frozen. Do not return used solution into the container. Keep cap tightly sealed when not in use. Discard 30 days after opening.

Conformance Statement

Mesa Laboratories Inc certifies that the above referenced product was tested using the N.I.S.T. traceable standards listed above, and meets or exceeds all published specifications as printed on the product label.



 Jamie Louie
 Director of Quality
 Instruments Division



 Ryan Rausch
 Production Manager