



# Certificate of Analysis



PJLA  
Testing  
Accreditation #66239

Mesa Laboratories Inc  
12100 West 6<sup>th</sup> Ave., Lakewood, CO 80228  
TEL (303) 987-8000 FAX (303) 987-8989  
[www.mesalabs.com](http://www.mesalabs.com)

Description: Combined Standard Solution **14.0 mS/cm 7.00 pH @ 25°C**  
Part Number: **02.0062**  
Lot No: **ML-P7C14-1442**  
Certification Date: **February 26, 2018**  
Expiration Date: **August 22, 2019**

Nominal Value (mS/cm)	Measured Value (mS/cm)	Tolerance @ 25°C (mS/cm)	Measured Value In Tolerance	Reference System Standard	Reference System Uncertainty @ 25°C (mS/cm)
<b>14.0000</b>	<b>13.999</b>	$\pm 0.03$ mS	Yes	NIST SRM 999	<b><math>\pm 0.008</math></b>

Nominal Value (pH Units)	Measured Value (pH Units)	Tolerance @ 25°C (pH Units)	Measured Value In Tolerance	Reference System Standards	Reference System Uncertainty @ 25°C (pH Units)
<b>7.000</b>	<b>7.000</b>	$\pm 0.03$	Yes	NIST SRM 186 NIST SRM 191	<b><math>\pm 0.006</math></b>

### Test Methods

- All analytical balances are calibrated by an ISO/IEC 17025:2005 accredited calibration laboratory. All balances are checked prior to use using an in-house procedure. Weights used for testing are traceable to NIST.
- All thermometers are NIST traceable through reference temperature probes that are calibrated by an ISO/IEC 17025:2005 accredited calibration laboratory.
- Conductivity measurements are taken at  $25^\circ\text{C} \pm 0.25^\circ\text{C}$  using SOP 900215-001 and are temperature compensated. Uncertainty is calculated using a 95% confidence interval using the Student's T Test for (n=30).
- pH measurements are taken at  $25^\circ\text{C} \pm 0.25^\circ\text{C}$  using EPA Method 150.1 and are temperature compensated. Uncertainty is calculated using an Expanded Uncertainty  $U=kuc$  (k=2.00).

### Intended Use

This standard solution is indicated for calibrating conductivity meters and pH meters.

### Hazardous Information

Please refer to the Material Safety Data Sheet available on our website for information on this material

### Stability and Storage

Protect from temperature extremes. Discard if solution has been frozen. Do not return used solution into the container. Keep cap tightly sealed when not in use. Discard 30 days after opening.

### Conformance Statement

Mesa Laboratories Inc certifies that the above referenced product was tested using the N.I.S.T. traceable standards listed above, and meets or exceeds all published specifications as printed on the product label. Our manufacturing laboratory is accredited to ISO/IEC 17025.

  
\_\_\_\_\_  
Jamie Louie  
Director of Quality  
Instruments Division

  
\_\_\_\_\_  
Ryan Rausch  
Production Manager