



Certificate of Analysis



Mesa Laboratories Inc
 12100 West 6th Ave., Lakewood, CO 80228
 TEL (303) 987-8000 FAX (303) 987-8989
www.mesalabs.com

Description: Combined Standard Solution **14.0 mS/cm 7.00 pH @ 25°C**
 Part Number: **02.0070**
 Lot No: **ML-P7C14-1467**
 Certification Date: **July 06, 2018**
 Expiration Date: **January 05, 2020**

Nominal Value (mS/cm)	Measured Value (mS/cm)	Tolerance @ 25°C (mS/cm)	Measured Value In Tolerance	Reference System Standard	Reference System Uncertainty @ 25°C (mS/cm)
14.0000	13.999	±0.03 mS	Yes	NIST SRM 999	±0.008

Nominal Value (pH Units)	Measured Value (pH Units)	Tolerance @ 25°C (pH Units)	Measured Value In Tolerance	Reference System Standards	Reference System Uncertainty @ 25°C (pH Units)
7.000	7.002	±0.03	Yes	NIST SRM 186 NIST SRM 191	±0.006

Test Methods

- All analytical balances are calibrated by an ISO/IEC 17025:2005 accredited calibration laboratory. All balances are checked prior to use using an in-house procedure. Weights used for testing are traceable to NIST.
- All thermometers are NIST traceable through reference temperature probes that are calibrated by an ISO/IEC 17025:2005 accredited calibration laboratory.
- Conductivity measurements are taken at 25°C ± 0.25°C using SOP 900215-001 and are temperature compensated. Uncertainty is calculated using a 95% confidence interval using the Student's T Test for (n=30).
- pH measurements are taken at 25°C ± 0.25°C using EPA Method 150.1 and are temperature compensated. Uncertainty is calculated using an Expanded Uncertainty U=kuc (k=2.00).

Intended Use

This standard solution is indicated for calibrating conductivity meters and pH meters.

Hazardous Information


Please refer to the Material Safety Data Sheet available on our website for information on this material

Stability and Storage

Protect from temperature extremes. Discard if solution has been frozen. Do not return used solution into the container. Keep cap tightly sealed when not in use. Discard 30 days after opening.

Conformance Statement

Mesa Laboratories Inc certifies that the above referenced product was tested using the N.I.S.T. traceable standards listed above, and meets or exceeds all published specifications as printed on the product label. Our manufacturing laboratory is accredited to ISO/IEC 17025.


 Jamie Louie
 Sr. Quality Manager
 Instruments Division


 Doug Weerstra
 Laboratory Manager